



Saponin Assay Kit

ARG83426 Saponin Assay Kit can be used to measure Saponin in tissue extracts and other biological fluids.

Catalog number: ARG83426

Package: 96 wells

For research use only. Not for use in diagnostic procedures.

TABLE OF CONTENTS

SECTION	Page
INTRODUCTION	3
PRINCIPLE OF THE ASSAY	3
MATERIALS PROVIDED & STORAGE INFORMATION	3
MATERIALS REQUIRED BUT NOT PROVIDED	4
TECHNICAL HINTS AND PRECAUTIONS	4
SAMPLE COLLECTION & STORAGE INFORMATION.....	4
REAGENT PREPARATION.....	5
ASSAY PROCEDURE.....	5
CALCULATION OF RESULTS	7
EXAMPLE OF TYPICAL RESULT	8

MANUFACTURED BY:

Arigo Biolaboratories Corporation

Address: 9F.-7, No. 12, Taiyuan 2nd St., Zhubei City,

Hsinchu County 302082, Taiwan

Tel: +886-3-6221320

Fax: +886-3-5530266

Email: info@arigobio.com

INTRODUCTION

Saponins, also selectively referred to as triterpene glycosides, are bitter-tasting usually toxic plant-derived organic chemicals that have a foamy quality when agitated in water. Saponins are widely distributed but found particularly in soapwort, a flowering plant, the soapbark tree and soybeans.

PRINCIPLE OF THE ASSAY

The Saponin Assay Kit can measure Saponin in tissue extracts and other biological fluids. The increase in absorbance at 540 nm is directly proportional to reactants of the reaction between substrate and Saponin.

MATERIALS PROVIDED & STORAGE INFORMATION

Store all components store at 4°C.

Component	Quantity	Storage
Microplate	1 X 96-well plate	
Standard	1 vial (lyophilized)	4°C
Assay Buffer	4 X 30 ml (ready to use)	4°C
Reagent Dye	1 vial (lyophilized)	4°C, keep in dark
Reagent Dye Diluent	5 ml	4°C
Plate Adhesive Strips	3 Strips	

MATERIALS REQUIRED BUT NOT PROVIDED

- Microplate reader capable of measuring absorbance at 540 nm
- Pipettes and pipette tips
- Deionized or distilled water
- Sulfuric acid (98%)- Analytical Reagent Grade
- Ice-water bath

TECHNICAL HINTS AND PRECAUTIONS

- Wear protective gloves, clothing, eye, and face protection especially while handling blood or body fluid samples.
- Store all components store at 4°C.
- Briefly spin down the reagents before use.
- It is highly recommended that the standards and samples be assayed in at least duplicates.
- Change pipette tips between the addition of different reagent or samples.

SAMPLE COLLECTION & STORAGE INFORMATION

The sample collection and storage conditions listed below are intended as general guidelines. Sample stability has not been evaluated.

Tissue - Weigh out 0.1 g tissue, homogenize with 1 ml Assay Buffer, then transfer it to the microcentrifuge tubes; incubate at 50 °C water bath for 1 hour; centrifuged at 10,000g for 10 minutes, take the supernatant into a new centrifuge tube for detection.

*Note: liquid samples can detect directly.

REAGENT PREPARATION

- **Standard:** Briefly centrifuge prior to opening. Add **1 ml** of **Assay Buffer** to dissolve standard, then add 0.1 ml into 0.1 ml **Assay Buffer**, the concentration will be 5 $\mu\text{mol /ml}$. Perform 2-fold serial dilution of the top standards to make the standard curve. The concentration of standard curve could be 5 $\mu\text{mol /ml}$, 2.5 $\mu\text{mol /ml}$, 1.25 $\mu\text{mol /ml}$, 0.63 $\mu\text{mol /ml}$, 0.31 $\mu\text{mol /ml}$, 0.16 $\mu\text{mol /ml}$, 0.08 $\mu\text{mol /ml}$, 0.04 $\mu\text{mol /ml}$.
- **Reagent Dye :** Add 5ml of **Reagent Dye Diluent** to dissolve Reagent Dye before use. Store at 4 °C in dark for 1 week after reconstitution.

ASSAY PROCEDURE

Standards and samples should be assayed in at least duplicates.

1. Sample wells: Add 50 μl of **Sample** into Sample wells.
2. Standard wells: Add 50 μl of **Standard** into Standard wells.
3. Blank wells: Add 50 μl **Assay Buffer** into Blank wells
4. Add 50 μl of **Reagent Dye** to each wells
5. Add 100 μl **Sulfuric acid** to down the inner wall slowly
6. Mix thoroughly while keeping in the ice bath.
7. Incubate at 60°C for 15 min.
8. Read the OD at 540 nm after cooling.

Saponin Assay Kit ARG83426

Summary of Saponin Assay Kit Procedure

Reagent	Sample	Standard	Blank
Sample	50 μ l	-	-
Standard	-	50 μ l	-
Assay Buffer	-	-	50 μ l
Reagent Dye	50 μ l	50 μ l	50 μ l
Sulfuric acid	100 μ l	100 μ l	100 μ l
Mix thoroughly while keeping in the ice bath			
Incubate at 60 °C for 15 minutes .			
Read the OD at 540 nm after cooling.			

Note:

1. Reagents must be added sequentially and should not be premixed prior to addition.
2. The concentrations can vary over a wide range depending on the different samples. For unknown samples, we recommend doing a pilot experiment & testing several doses to ensure the readings are within the standard curve range.
3. For colored samples, we recommend setting a parallel sample background control well with same volume of sample only. Other reagents were replaced by distilled water to the same total volume. Subtract the OD value of the sample background control from the OD value of the sample to correct for interference from the sample's own color.

CALCULATION OF RESULTS

1. Calculate the average absorbance values for each set of samples, standard and blank.

2. Calculation:

A. Definition:

C_{Standard} : the standard concentration, $5 \mu\text{mol} / \text{ml} = 0.005 \text{ mmol} / \text{ml}$;

W : the weight of sample, g;

V_{Sample} : the volume of sample in assay procedure, $50 \mu\text{l} = 0.05 \text{ ml}$;

V_{standard} : the volume of standard in assay procedure, $50 \mu\text{l} = 0.05 \text{ ml}$;

V_{assay} : the volume of Assay Buffer, $1000 \mu\text{l} = 1 \text{ ml}$.

B. Formula:

1. According to the slope of the standard curve

Calculate the sample concentration in ASSAY PROCEDURE according to the slope of the standard curve

Saponin (mmol/L) =

$$[(\text{OD}_{\text{Sample}} - \text{OD}_{\text{Blank}}) - \text{Intercept}] / (\text{Slope}) \times n$$

Calculate the initial concentration according to sample preparation procedure.

2. According to one point of the standard OD and concentration a).

According to the weight of sample

Saponin ($\mu\text{mol/g}$) =

$$[(C_{\text{Standard}} \times V_{\text{standard}}) \times (\text{OD}_{\text{Sample}} - \text{OD}_{\text{blank}})] / [(\text{OD}_{\text{Standard}} - \text{OD}_{\text{Blank}}) \times (W \times V_{\text{Sample}} / V_{\text{assay}})]$$

$$= 0.005 \times (\text{OD}_{\text{Sample}} - \text{OD}_{\text{blank}}) / [(\text{OD}_{\text{Standard}} - \text{OD}_{\text{Blank}}) \times W]$$

b). According to the volume of sample

Saponin ($\mu\text{mol/ml}$) =

$$\frac{[(C_{\text{Standard}} \times V_{\text{standard}}) \times (OD_{\text{Sample}} - OD_{\text{blank}})]}{[(OD_{\text{Standard}} - OD_{\text{Blank}}) \times V_{\text{Sample}}]} \\ = 0.005 \times (OD_{\text{Sample}} - OD_{\text{blank}}) / (OD_{\text{Standard}} - OD_{\text{Blank}})$$

3. Detection range:

The detection range is from $0.04 \mu\text{mol/ml}$ - $5 \mu\text{mol/ml}$.

4. If the samples have been diluted, the calculated activity must be further converted by the appropriate dilution factor according to the sample preparation procedure as described above.

EXAMPLE OF TYPICAL RESULT

The following data is for demonstration only and cannot be used in place of data generations at the time of assay. Please note this data is for demonstration only and serially diluted standards are necessary for this kit.

